## **Technical Datasheet**

Date: 11.11.16 Page 1 of 1



DIN EN ISO 13485 certified

		EN ISO certified
Article - No:	08-552-0005	Revision: 1
Remarks	The survival rate of bacteria corresponds to the standard: DIN 58942-4 The filling amount of the liquid to be examined must be 10 ml, with a maximum tolerance of +/- 10%	single use only

## **Specification**

Product	EXACTOBAC-L®, 10ml, milk specimen tube PP, Ø16x100 mm, with fitted screw cap PE, with lyophilised bacteriostat, tube transparent, cap blue, graduated, CE/IVD	
Classification	CE/IVD	
Dimensions	outer diameter: 16,00 mm (+/- 0,2 mm)	
	height (withour screw cap): 100,00 mm (+/- 0,2 mm)	
Colour	tube transparent, cap blue	
Graduation	graduated	
Nominal Capacity	10ml	
Material Characteristics	tube: polypropylene (PP)	
	cap: polyethylene (PE)	
Product Characteristics	biologically inert	
	physiologically harmless	
	media: lyophilisate on boric acid	
	sturdy, shock resistant	
	high chemical resistance	
	transparent tube and yellow cap	
Shelf - Life	60 months at controlled storage conditions	
Packaging	pcs per bag: 1.000	
	pcs per carton: 1.000	
	pcs per pallet: 24.000	
Minimum order qty.	1.000 pcs	
Storage	in dry condition: 40% - 60% rel. humidity	
	bags and cartons originally closed	
	storage temperature +10°C to maximal +30°C	
	pallets cannot be stacked	
Approved by nerbe plus	quality department	







