

Product Specification

SAPHIR PEARL

Disposable Nitrile Examination Gloves

- Powder free
- Lilac
- Chemical resistance protective gloves
- Latex-free
- Suitable for handling food
- Non-sterile



SAFETY MEETS QUALITY

CE 0197



Material 100% Synthetic Nitrile Latex

Performance requirements

Characteristics	Inspection Level	Acceptable Quality Level
Dimensions	S-2	4.0
Physical Properties	S-2	4.0
Freedom from Holes	G-1	1.5
Major Visual defects	G-1	1.5
Minor Visual defects	G-1	4.0

Dimensions

Number	6511	6512	6513	6514	6515
Size	XS = 5-6	S = 6-7	M = 7-8	L = 8-9	XL = 9-10
Width	75 ± 5 mm	80 ± 10 mm	95 ± 10 mm	110 ± 10 mm	≥ 110 mm
Length	≥ 240 mm	≥ 240 mm	≥ 240 mm	≥ 240 mm	≥ 240 mm

Thickness

Measuring Points	Single Wall (mm)
Cuff (distance 25 mm to the beaded edge)	0.08 ± 0.03
Palm (in the middle of the palm)	0.10 ± 0.03
Fingertip (distance 15 mm to top of middle finger)	0.12 ± 0.03

Physical Properties

Standard	Before Aging	After Aging
Median force at break (N)	min. 6.0	min. 6.0
Ultimate Elongation (%)	min. 500	min. 400
Tensile strength (MPa)	min. 14.0	min. 14.0

Powder No powder lubricant added

Protein content No proteins contained

Cuff Beaded cuff

Design Ambidextrous

Surface Micro-roughened finger tips

Color Lilac

Quality assurance EG-construction sample test TÜV Rheinland – CE 0197

PSA – EN 374 CAT III, EN 388, EN 420, 89/686/EEC
 MPG – EN 455 Teil 1-4
 EG-Standard 596/2009
 EG-Standard 202/2014
 EN ISO 9001:2008
 EN ISO 13485:2012

Packing Style 100 pcs per dispenser, 10 dispensers per carton

Hint The above information apply to the standard version.
 Subject to change.

HUBERLAB. AG
 Industriestrasse 123
 4147 Aesch

T 061 717 99 77
 F 061 711 93 42

www.huberlab.ch
 info@huberlab.ch

HUBERLAB.

committed to science